

***International Intercomparison Exercise on Natural Radiation Measurements  
under Field Conditions***

***Saelices el Chico (Salamanca, Spain), May 23-27, 2011***

***Measurement of External Environmental Gamma Radiation***

**J.C. Sáez Vergara, E. Correa Garcés  
CIEMAT, Madrid (Spain)**

- External environmental dose rate is probably the most used radiological parameter in radiation protection.
- It describes the exposure to external radiation sources, which dominates most of the human practices and is the second in natural sources after radon inhalation.
- In principle, it is easy to measure , standards and calibration preocedures are well established, many instruments are available at reasonable costs, large experience since radioactivity discovery...

BUT ...



# EURADOS: Comparison of early warning instruments since 1999

| Country         | Responsible for national network<br>( <u>Intercomparison</u> )           | Other participants<br>( <u>Intercomparison</u> )                        |
|-----------------|--------------------------------------------------------------------------|-------------------------------------------------------------------------|
| AUSTRIA         | Federal Chancellery, Vienna (1999)                                       | <u>Bitt Technology</u> (1999)                                           |
| CZECH REPUBLIC  | National Radiation Protection Institute,<br>NRPI (1999)                  |                                                                         |
| DENMARK         | <u>Risø National Laboratory</u> , RNL (1999)                             |                                                                         |
| FRANCE          | Institute de Radioprotection et Sureté<br><u>Nucleaire</u> , IRSN (2002) |                                                                         |
| GERMANY         | <u>Bundesamt für Strahlenschutz</u> , BfS<br>(1999)                      | <u>Physikalisch-Technische<br/>Bundesanstalt</u> , PTB<br>(1999 & 2002) |
| GREECE          | Greek Atomic Energy <u>Commision</u> , GAEC<br>(2002)                    | Aristotle University of<br>Thessaloniki (2002)                          |
| HUNGARY         | <u>Paks Nuclear Power Plant</u> (2002)                                   | Atomic Energy Research<br>Institute (2002)                              |
| THE NETHERLANDS | National Institute of Public Health and the<br>Environment, RIVM (1999)  |                                                                         |
| PORTUGAL        | <u>Direcção Geral Ambiente</u> , DGA (1999)                              |                                                                         |
| SPAIN           | Consejo de Seguridad Nuclear, CSN (1999)                                 | CIEMAT (1999 & 2002)                                                    |
| SWEDEN          | Swedish Radiation Protection Authority,<br>SSI (2002)                    |                                                                         |
| SWITZERLAND     | Swiss Nuclear Safety Inspectorate, HSK<br>(2002)                         |                                                                         |
|                 | Institut de <u>Radiophysique Appliquée</u> , IRA<br>(2002)               |                                                                         |
| UNITED KINGDOM  | No participant                                                           | Stirling University (2002)                                              |
|                 |                                                                          | Consultant (1999 & 2002)                                                |

# INSTRUMENTS PARTICIPATING IN EURADOS 1999


| ID | Type  | Application | Manufacturer  | Quantity | Units | Home Source              |
|----|-------|-------------|---------------|----------|-------|--------------------------|
| 11 | PC    | Network     | Bitt          | H*(10)   | nSv/h | Cs137                    |
| 12 | PC    | Other       | Bitt          | H*(10)   | nSv/h | Cs137                    |
| 21 | PC    | Other       | FAG           | H*(10)   | nSv/h | Cs137                    |
| 22 | PC    | Other       | Bitt          | H*(10)   | nSv/h | Cs137                    |
| 23 | PC    | Network     | Berthold      | H*(10)   | nSv/h | Cs137                    |
| 24 | GM    | Other       | Berthold      | H*(10)   | nSv/h | Cs137                    |
| 25 | IC    | Other       | Reuter Stokes | X        | μR/h  | Cs137                    |
| 26 | NaISc | Other       | Tesla         | Kair     | nGy/h | Cs137                    |
| 31 | IC    | Other       | Reuter Stokes | X        | μR/h  | Information not supplied |
| 32 | IC    | Network     | Reuter Stokes | X        | μR/h  | Information not supplied |
| 33 | NaISc | Network     | Bicron        | counting | cps   | Cs137                    |
| 41 | GM    | Other       | Hormann       | Hx(10)   | nSv/h | Information not supplied |
| 42 | GM    | Network     | Hormann       | Kair     | nGy/h | Information not supplied |
| 43 | GM    | Other       | Hormann       | Kair     | nGy/h | Information not supplied |
| 51 | PC    | Network     | Bitt          | H*(10)   | nSv/h | Co60                     |
| 52 | PC    | Other       | Bitt          | H*(10)   | nSv/h | Cs137                    |
| 53 | GM    | Other       | Genitron      | H*(10)   | nSv/h | Cs137                    |
| 61 | GM    | Network     | Hormann       | Kair     | nGy/h | Cs137                    |
| 71 | PSc   | Other       | MAB           | H*(10)   | nSv/h | Cs137                    |
| 72 | GM    | Network     | Berthold      | H*(10)   | nSv/h | Cs137                    |
| 73 | GM    | Other       | Genitron      | Kair     | nGy/h | Cs137                    |
| 74 | IC    | Other       | Reuter Stokes | X        | μR/h  | Cs137                    |




| ID  | Type | Application | Manufacturer  | Quantity | Units | Home Source    |
|-----|------|-------------|---------------|----------|-------|----------------|
| 081 | GM   | Network     | Rados         | H*(10)   | nSv/h | Cs137 and Co60 |
| 082 | IC   | Other       | Reuter Stokes | H*(10)   | nSv/h | Cs137 and Co60 |
| 083 | PSc  | Other       | APVL          | H*(10)   | nSv/h | Cs137          |
| 091 | IC   | Network     | Reuter Stokes | Dair     | nGy/h | Cs137          |
| 101 | PC   | Network     | Bitt          | H*(10)   | nSv/h | Cs137          |
| 102 | PC   | Other       | Bitt          | H*(10)   | nSv/h | Cs137          |
| 111 | PC   | Other       | Eberline      | H*(10)   | nSv/h | Cs137          |
| 112 | GM   | Other       | Rados         | H*(10)   | nSv/h | Cs137          |
| 113 | IC   | Network     | SSI           | H*(10)   | nSv/h | Cs137          |
| 121 | GM   | Network     | Horman        | H*(10)   | nSv/h | Cs137          |
| 122 | IC   | Other       | Reuter Stokes | H*(10)   | nSv/h | Cs137          |
| 123 | GM   | Network     | Technidata    | H*(10)   | nSv/h | Cs137          |
| 131 | IC   | Other       | Reuter Stokes | X        | μR/h  | Co60           |
| 132 | IC   | Other       | Reuter Stokes | X        | μR/h  | Co60           |

- ☆ Most of the instruments measures H\*(10)
- ☆ Instruments are mostly calibrated using Cs137 sources
- ☆ Two instruments use some background value which are automatically subtracted from the current measurement


- Automatic device to program controlled gamma plume profiles
- Two  $^{137}\text{Cs}$  encapsulated sources which yield dose increments from 3% to 35%.
- Step time length: 10 minutes
- Plume simulation was produced continuously for 20 hours




- Automatic device to program controlled gamma plume profiles kindly provided by Riso National Laboratory
- $^{137}\text{Cs}$  and  $^{60}\text{Co}$  encapsulated sources yielding dose increments from 7% to 45%.
- Step time length: 10 minutes
- Plume simulation was produced continuously for 20 hours






(a) Instrumentos calibrados en términos de Dosis equivalente ambiental, H\*(10)



# EURADOS 2002: Comparison of early warning instruments



CONCLUSION: Relevant differences often occurs when comparing simultaneous environmental dose rate measurements!!



Why? Because accurate dose rate estimation is not so simply as:


$$\text{Dose Rate} = \text{Signal} \times \text{Calibration Factor}$$

In fact, these are probably the most relevant factors affecting estimations:

- Quantity (and units!!!).
- Calibration conditions (procedure, reference energies, ...)
- Sensitivity (time resolution).
- Inherent background.
- Photon Energy response.
- Cosmic Response.
- Angular Response.
- Environmental response (temperature, humidity, light...)



# Quantities for external radiation dosimetry






- Absorbed Dose and Kerma are numerically equal in electronic equilibrium conditions:

$$D_m \cong K_m$$



- Air Kerma and Exposure relationship is almost independent of photon energy.

$$X = K_{\text{aire}} \cdot (1 - g_{\text{aire}}) \cdot \left( \frac{e}{W} \right)$$

- Conversion Factors are needed to relate Air Kerma and Equivalent Dose quantities.




- $^{137}\text{Cs}$ ,  $^{60}\text{Co}$  and  $^{226}\text{Ra}$  encapsulated sources
- Reference distance: 5 m
- $H^*(10)$  rates: 330-400 nSv/h
- Reference values calculated by Reference instrument and MonteCarlo methods



# Results: Calibration Check

## Background at the PTB Free Field




**84.0±5.9 nSv/h (7%)**

**84.5±5.0 nSv/h (6%)**

**85.7±6.6 nSv/h (8%)**

**85.5±6.1 nSv/h (7%)**

## Source exposures at the PTB Free Field






**Co60: 520±48 nSv/h (9%)**

**Ra 226: 496±35 nSv/h (7%)**


**Cs137: 433±13 nSv/h (3%)**

NOTE: Experimental inherent background has been substracted




- Important over-response figures for GM and PC when dealing with <sup>60</sup>Co and <sup>226</sup>Ra
- Slight but consistent differences have been encountered for Reuter-Stokes HPIC


- Located 925 m depth in the Asse Salt Mine (Germany)
- Rock Salt Activity (Bq/kg): 2-4 ( $^{40}\text{K}$ ),  $<0.1$  ( $^{238}\text{U}$ ),  $<0.01$  ( $^{232}\text{Th}$ )
- $H^*(10)$  dose rate:  $\approx 1$  nSv/h
- Excellent room conditions and irradiation facilities (collimated beams)





- $^{241}\text{Am}$ ,  $^{57}\text{Co}$ ,  $^{137}\text{Cs}$ ,  $^{226}\text{Ra}$  and  $^{60}\text{Co}$  encapsulated sources (60-1300 keV)
- Reference distance: 2 m
- Kair rates: 30-130 nGy/h
- Inherent background was obtained after overnight exposure without any radioactive source.



# Measurement of the cosmic radiation






## Cosmic&Terrestrial components (2/2)


### Reference levels for cosmic & terrestrial

- Muon dose rate at PTB-Braunschweig (PTB coincidence PC instrument):  
**34 nSv/h**
- Soft cosmic component ( $e^-$ , photons) is 30-40% of muon contribution (fluence):  
**10-13 nSv/h**  
totalising  
**44-47 nSv/h**
- Dose rate at the platform from the Reference instrument 131/IC:  
**44 nSv/h**
- Reference level for terrestrial component (by difference Lakeshore-Platform):  
**21 nSv/h**




Cosmic response relative to 131/IC (44 nSv/h)



Terrestrial response relative to 131/IC (21 nSv/h)






# In situ Gamma Spectrometry






## Reference levels for spectrometers

- Estimate of point source activity:  
Nominal source activity
- Other field measurements (no sources):  
Homogeneous nuclide distribution,  
Semi-infinite half space
- Air kerma rate from the direct beam  
(scattered radiation is not included)
- Values based on reference instrument  
(131/IC) measurements




# In situ gamma spectrometry



# In situ gamma spectrometry



# In situ gamma spectrometry



**Dose Rate = f(Signal, Quantity, Calibration, BKGND, Cosmic, Photon Energy, Angular, others)**

**You should read the indications of your dose rate monitor and then apply the corresponding corection factors according to the field conditions.**



GOBIERNO  
DE ESPAÑA

MINISTERIO  
DE CIENCIA  
E INNOVACIÓN

**Ciemat**  
Centro de Investigaciones  
Energéticas, Medioambientales  
y Tecnológicas

## ***Realistic calibration fields for environmental Dose rates and in situ gamma spectrometry***



***Wismuth GMBH, Gera, Germany***

# Cosmic & Terrestrial Studies in ... Santander!


Resultados GENIE-2000/ISOCS ordenados por series

|         | Santander CMT<br>Bq/kg | Santander Bahía<br>Bq/kg | Aguilar Orilla<br>Bq/kg | Aguilar Lago<br>Bq/kg |
|---------|------------------------|--------------------------|-------------------------|-----------------------|
| Th-234  | <AMD                   | <AMD                     | <AMD                    | <AMD                  |
| Pa-234m | <AMD                   | <AMD                     | <AMD                    | <AMD                  |
| Ra-226  | 79.3 ± 14.8            | 19.5 ± 5.5               | 28.9 ± 15.7             | 17.0 ± 8.4            |
| Pb-214  | 34.1 ± 1.0             | 4.1 ± 0.4                | 21.1 ± 1.0              | 7.3 ± 0.5             |
| Bi-214  | 37.1 ± 0.8             | 3.3 ± 0.2                | 22.8 ± 0.9              | 6.2 ± 0.5             |
| Pb-210  | <AMD                   | <AMD                     | <AMD                    | <AMD                  |
| Ac-228  | 34.0 ± 1.0             | <AMD                     | 17.4 ± 0.9              | 3.8 ± 0.4             |
| Pb-212  | 31.7 ± 0.9             | 1.8 ± 0.4                | 17.1 ± 0.7              | 7.7 ± 0.6             |
| Bi-212  | 31.3 ± 4.3             | 1.3 ± 1.2                | 22.2 ± 4.3              | <AMD                  |
| Tl-208  | 13.3 ± 0.3             | 0.7 ± 0.1                | 6.9 ± 0.4               | 1.0 ± 0.1             |
| K-40    | 391.7 ± 9.7            | 28.8 ± 1.6               | 383.4 ± 9.8             | 14.1 ± 1.6            |
| Cs-137  | 0.4 ± 0.3              | 0.0 ± 0.0                | 8.8 ± 0.5               | 0.0 ± 0.0             |

Resumen de tasa de kerma en aire

|                    | Santander CMT<br>43° 29' N, 3° 48W<br>58 m<br>nGy/h | Santander Bahía<br>43° 29' N, 3° 48W<br>0 m<br>nGy/h | Aguilar Orilla<br>42° 47' N, 4° 2' W<br>970 m<br>nGy/h | Aguilar Lago<br>42° 47' N, 4° 2' W<br>970 m<br>nGy/h |
|--------------------|-----------------------------------------------------|------------------------------------------------------|--------------------------------------------------------|------------------------------------------------------|
| Serie U-238        | 16.8 ± 0.4                                          | 1.6 ± 0.1                                            | 10.3 ± 0.4                                             | 2.9 ± 0.2                                            |
| Serie Th-232       | 13.6 ± 0.5                                          | 0.3 ± 0.1                                            | 7.2 ± 0.5                                              | 1.4 ± 0.1                                            |
| K-40               | 16.3 ± 0.4                                          | 1.2 ± 0.1                                            | 16.0 ± 0.4                                             | 0.6 ± 0.1                                            |
| Cs-137             | 0.1 ± 0.1                                           | 0.0 ± 0.0                                            | 1.5 ± 0.1                                              | 0.0 ± 0.0                                            |
| Terrestre SEGIS    | 46.8 ± 1.4                                          | 3.1 ± 0.3                                            | 34.9 ± 1.4                                             | 4.9 ± 0.4                                            |
| Cósmica SEGIS      | 36.5 ± 2.8                                          | 37.1 ± 2.6                                           | 54.3 ± 3.6                                             | 55.4 ± 3.6                                           |
| Cósmica CARI6      | 36.9                                                | 35.9                                                 | 53.0                                                   | 53.0                                                 |
| Total SEGIS        | 83.2 ± 3.1                                          | 40.2 ± 2.6                                           | 89.3 ± 3.8                                             | 60.2 ± 3.6                                           |
| Total PIC          | 79.1 ± 0.9                                          | 40.1 ± 0.9                                           | 89.4 ± 0.9                                             | 52.8 ± 0.9                                           |
| Cociente SEGIS/PIC | 1.05 ± 0.05                                         | 1.00 ± 0.09                                          | 1.00 ± 0.05                                            | 1.14 ± 0.09                                          |







# 2011 Intercomparison at Saelices el Chico

|                | <b>Station 1</b><br><b>Bq/kg</b> |   |      | <b>Point 17</b><br><b>Bq/kg</b> |   |     | <b>Station 2</b><br><b>Bq/kg</b> |   |      |
|----------------|----------------------------------|---|------|---------------------------------|---|-----|----------------------------------|---|------|
| <b>K-40</b>    | 750                              | - | 1150 | 600                             | - | 900 | 750                              | - | 1200 |
| <b>Th-234</b>  | < AMD                            |   |      | 250                             | - | 350 | < AMD                            |   |      |
| <b>Pa-234m</b> | < AMD                            |   |      | 6                               | - | 7   | < AMD                            |   |      |
| <b>Ra-226</b>  | < AMD                            |   |      | 700                             | - | 800 | 5500                             | - | 7500 |
| <b>Pb-214</b>  | 30                               | - | 40   | 220                             | - | 230 | 4500                             | - | 5500 |
| <b>Bi-214</b>  | 30                               | - | 50   | 210                             | - | 250 | 4000                             | - | 6000 |
| <b>Pb-210</b>  | < AMD                            |   |      | < AMD                           |   |     | 2500                             | - | 9500 |
| <b>Ac-228</b>  | 15                               | - | 30   | 30                              | - | 40  | 40                               | - | 80   |
| <b>Pb-212</b>  | 15                               | - | 20   | 30                              | - | 40  | 50                               | - | 100  |
| <b>Bi-212</b>  | 25                               | - | 30   | 50                              | - | 80  | < AMD                            |   |      |
| <b>Tl-208</b>  | 5                                | - | 10   | 15                              | - | 20  | 10                               | - | 30   |
| <b>U-235</b>   | < AMD                            |   |      | < AMD                           |   |     | < AMD                            |   |      |
| <b>Th-227</b>  | < AMD                            |   |      | < AMD                           |   |     | 1300                             | - | 2300 |
| <b>Ra-223</b>  | < AMD                            |   |      | < AMD                           |   |     | 1400                             | - | 2300 |
| <b>Cs-137</b>  | < AMD                            |   |      | < AMD                           |   |     | < AMD                            |   |      |

# 2011 Intercomparison at Saelices el Chico

|         | Zona<br>Punto de medida   | Estación 1<br>1 | Estación 1<br>2 | Estación 1<br>4 | Estación 1<br>6 | Estación 1<br>8 | Estación 1<br>Promedio<br>Puntos 1,2,4,6,8 | Estación 1<br>17 |
|---------|---------------------------|-----------------|-----------------|-----------------|-----------------|-----------------|--------------------------------------------|------------------|
|         | X -UTM, m                 | 701704.7        | 701707.3        | 701708.5        | 701700.9        | 701700.8        |                                            | 701696.1         |
|         | Y-UTM, m                  | 4501191.2       | 4501196.1       | 4501187.0       | 4501186.5       | 4501195.6       |                                            | 4501238.9        |
|         | Z, m (snm)                | 719.3           | 719.1           | 719.1           | 720.0           | 719.1           |                                            | 719.8            |
|         | Notas                     |                 |                 |                 |                 |                 |                                            |                  |
| CIP     | Fecha de medida           | 25/04/2011      | 25/04/2011      | 25/04/2011      | 25/04/2011      | 25/04/2011      | 25/04/2011                                 | 25/04/2011       |
| RSS-112 | K <sub>aire</sub> , nGy/h | 111 ± 4         | 112 ± 4         | 110 ± 5         | 110 ± 4         | 107 ± 4         | 110 ± 2                                    | 173 ± 5          |
| RSS-131 | K <sub>aire</sub> , nGy/h | 112 ± 4         | 113 ± 5         | 110 ± 4         | 111 ± 4         | 105 ± 4         | 110 ± 3                                    | 174 ± 5          |

|         | Zona<br>Punto de medida   | Estación 2<br>20 | Estación 2<br>21 | Estación 2<br>22 | Estación 2<br>23 | Estación 2<br>24 | Estación 2<br>Promedio<br>Puntos 20 a 24 |
|---------|---------------------------|------------------|------------------|------------------|------------------|------------------|------------------------------------------|
|         | X -UTM, m                 | 701541.8         | 701539.6         | 701537.7         | 701542.8         | 701545.6         |                                          |
|         | Y-UTM, m                  | 4502440.1        | 4502436.4        | 4502440.6        | 4502442.9        | 4502439.0        |                                          |
|         | Z, m (snm)                | 719.2            | 719.5            | 720.1            | 720.0            | 719.3            |                                          |
|         | Notas                     |                  |                  |                  |                  |                  |                                          |
| CIP     | Fecha de medida           | 26/04/2011       | 26/04/2011       | 26/04/2011       | 26/04/2011       | 26/04/2011       | 26/04/2011                               |
| RSS-112 | K <sub>aire</sub> , nGy/h | 1862 ± 12        | 1831 ± 15        | 2008 ± 17        | 1901 ± 12        | 1815 ± 28        | 1883 ± 69                                |
| RSS-131 | K <sub>aire</sub> , nGy/h | 1803 ± 66        | 1785 ± 67        | 2005 ± 31        | 1908 ± 30        | 1798 ± 68        | 1860 ± 85                                |

**Station 1 (Points 1 a 9):**  
**Point 17:**  
**Station 2 (Pointss 20 a 24):**

**0.110 ± 0.005 µGy/h (k=2)**  
**0.173 ± 0.005 µGy/h (k=2)**  
**1.80 ± 0.05 µGy/h (k=2)**



- In each station, please measure as many points as you can **WITHIN** the signals.
- **1.00 m** high are indicated in some points.
- Results should be given in terms of air kerma. Conversion factors are given below.
- You should send a form per instrument, filled with all the requested information. **PLEASE DON'T FORGET TO EXPRESS YOUR UNCERTAINTY.**

| Nuclide (Mean Photon Energy) | $H^*(10)/K_{\text{air}}$<br>Sv/Gy | $H^*(10)/X$<br>nSv/ $\mu$ R | $K_{\text{air}}/X$<br>nGy/ $\mu$ R |
|------------------------------|-----------------------------------|-----------------------------|------------------------------------|
| $^{241}\text{Am}$ (59.5 keV) | 1.740                             | 15.250                      | 8.764                              |
| $^{57}\text{Co}$ (122 keV)   | 1.447                             | 12.683                      | 8.764                              |
| $^{137}\text{Cs}$ (661 keV)  | 1.200                             | 10.517                      | 8.764                              |
| $^{226}\text{Ra}$ (840 keV)  | 1.179                             | 10.336                      | 8.764                              |
| $^{60}\text{Co}$ (1.25 MeV)  | 1.160                             | 10.166                      | 8.764                              |



## INTERNATIONAL INTERCOMPARISON EXERCISE ON NATURAL RADIATION MEASUREMENTS UNDER FIELD CONDITIONS

*Saelices el Chico (Salamanca), Spain May 23-27, 2011*

### *External Gamma Radiation Exercise*

Participant ID:

#### INSTRUMENT DESCRIPTION:

Instrument Model:

Manufacturer:

Detector Type: Ion Chamber, GM, Proportional Counter, Scintillator, other (please specify)

CALIBRATION (according to an existing calibration certificate from a metrological laboratory or from the manufacturer)

Quantity: Exposure (X), Air Kerma (K<sub>air</sub>), Air Absorbed Dose (D<sub>air</sub>), Ambient Dose Equivalent H\*(10), Photon Dose Equivalent H<sub>p</sub>(10), other (please specify):

Nuclide and Photon Energy: Cs-137, Co-60, Ra-226, other (please specify)

If necessary, please provide the conversion factors from instrument readings to Air Kerma rate at 661 keV/photons (Cs-137) in nGy/h:

| Point # | INSTRUMENT READINGS |                  |       | AIR KERMA RATE, nGy/h |                  |
|---------|---------------------|------------------|-------|-----------------------|------------------|
|         | Value               | Uncertainty(k=1) | Units | Value                 | Uncertainty(k=1) |
| 1       |                     |                  |       | 0                     | 0                |
| 2       |                     |                  |       | 0                     | 0                |
| 4       |                     |                  |       | 0                     | 0                |
| 6       |                     |                  |       | 0                     | 0                |
| 8       |                     |                  |       | 0                     | 0                |
| 17      |                     |                  |       | 0                     | 0                |
| 20      |                     |                  |       | 0                     | 0                |
| 21      |                     |                  |       | 0                     | 0                |
| 22      |                     |                  |       | 0                     | 0                |
| 23      |                     |                  |       | 0                     | 0                |
| 24      |                     |                  |       | 0                     | 0                |

Other comments from the participant:

Date:

Signature: